


Twice Daily Vaccine Fridge Temperature Chart

(for both domestic and purpose-built models)



The acceptable temperature range is + 2°C to + 8°C but **Strive for five!** (+ 5°C)

Please record current, minimum and maximum temperatures twice daily then reset your thermometer.

MONTH..... YEAR

Date	Temperature * - Morning			Comm ents Eg VD	Initials	Temperature * - Afternoon			Comm ents Eg BC	Initials
	Current	HI	LOW			Current	HI	LOW		
1										
2										
3										
4										
5										
6										
7										
8										
9										
10										
11										
12										
13										
14										
15										
16										
17										
18										
19										
20										
21										
22										
23										
24										
25										
26										
27										
28										
29										
30										
31										

Abbreviations:

VO - Vaccines ordered

FD - Freezer defrosted

TTD - Thermostat turned down

WE - Weekend

VD - Vaccine delivered

TTU - Thermostat turn up

TC - Thermometer Calibrated

PH - Public holiday

DO - Door opened for extended time

CF - Cleaning of inside of fridge

BC - Battery on thermometer changed

IMPORTANT

- If fridge temperature is below +2°C or above +8°C, contact Public Health Unit asap. Do not use these vaccines until you have received advice from PHU. The exception is a brief rise in temperature up to 12°C for less than 15 minutes associated with opening the fridge door.
- Data loggers are compulsory. Download and review the data from the logger weekly.

Northern Sydney Public Health Unit
1300 066 055 or fax 02 9477 9019

Sydney North PHN 02 9432 8250

NSW Vaccine Centre 1300 656 132